



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 1 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Attenuation (10 MHz to 18 GHz)	Using Power Sensors with Signal Generators by Substitution Method	0.5 dB to 50 dB	0.37 dB to 0.87 dB
2	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Attenuation (9 kHz to 10 MHz)	Using Power Sensors with Signal Generators by direct Method	0.5 dB to 33 dB	0.33 dB to 0.39 dB
3	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Frequency / Generators	Using Rubidium Source & Frequency Counter by Direct Method	9 kHz to 14 GHz	1.3 Hz to 19.4 Hz
4	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Frequency/Generators	Using Rubidium Source & Frequency Counter at Single Frequency by Direct Method	10 MHz	0.081 Hz



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 2 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Power / Generators (10 MHz to 18 GHz)	Using Power Sensors by Direct Method	10 dBm to (-) 40 dBm	0.41 dB to 0.64 dB
6	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Power / Generators (9 kHz to 10 MHz)	Using Power Sensors by Direct Method	13 dBm to (-) 20 dBm	0.36 dB to 0.41 dB
7	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Electric Field / Electromagnetic Field Sensor & Probe (9 kHz to 18000 MHz)	Using Electric Field Isotropic Probe & Electric Field Meter by Absolute Method (IEEE-1309-2013) - Type B, By TEC/SD/DD/CAL-EMF/01/FEB-19, By IEC 61000-4-3(2020) Annexure K	2 V/m to 100 V/m	14 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 3 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
8	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Electric Field/ Electromagnetic Field Sensor & Probe (80 MHz to 1000 MHz)	Using Electric Field Isotropic Probe & Electric Field Meter by Absolute Method (IEEE-1309-2013) - Type B, By TEC/SD/DD/CAL-EMF/01/FEB-19, By IEC 61000-4-3(2020) Annexure K	2 V/m to 500 V/m	14 %
9	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Attenuation (9 kHz to 18 GHz)	Using Power Sensors with Signal Generators by Direct Method	0.5 dB to 50 dB	0.1 dB to 0.65 dB
10	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Frequency / Receivers	Using Signal Generator, Reference Frequency Standard & Frequency Counter by Comparison / Reference Method	9 kHz to 18 GHz	1.3 Hz to 78 Hz
11	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Power / Receivers (9 kHz to 10 MHz)	Using Power Sensors with Signal Generators & Power Sensors by Comparison/ Substitution Method	10 dBm to (-) 20 dBm	0.43 dB to 0.47 dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11,
SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2765

Page No

4 of 9

Validity

25/06/2022 to 24/06/2024

Last Amended on

30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
12	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Power/ Receivers (10 MHz to 18 GHz)	Using Power Sensors with Signal Generators & Power Sensors by Comparison/ Substitution Method	+10 dBm to (-) 40 dBm	0.47 dB to 0.70 dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 5 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per ANSI C 63.5: 2017, CISPR 16-1-6: 2014 + Amd 1: 2017 + Amd 2: 2022	1 GHz to 18 GHz	2dB
2	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per SAE ARP 958 Rev. D	1 GHz to 18 GHz	2dB
3	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	SAE ARP 958 Rev. E	1 GHz to 18 GHz	-2 dB to 2 dB
4	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	SAE ARP 958 Rev. E	20 MHz to 1 GHz	-2 dB to 2 dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 6 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per ANSI C 63.5: 2017, CISPR 16-1-6: 2014 + Amd 1: 2017 + Amd 2: 202	30 MHz to 1 GHz	1.58dB
6	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per SAE ARP 958 Rev. D	30 MHz to 1 GHz	2dB
7	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Factor (AF)	Using Vector Network Analyzer as per CISPR 16-1-6: 2014 + Amd 1: 2017 + Amd 2: 2022, ANSI C63.5: 2017	9 kHz to 30 MHz	1.8dB
8	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Pair Reference site attenuation (Aapr)	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per CISPR 16-1-4: 2019 Clause 6.6.4, 6.10.2	30 MHz to 18 GHz	1dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 7 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
9	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Return Loss	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per CISPR 16-1-6: 2014 + Amd 1: 2017 + Amd 2: 2022: Section A.8.7	30 MHz to 18 GHz	2.1 dB
10	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Symmetry	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per ANSI C 63.5-2017: Section 4.4.3	30 MHz to 300 MHz	2 dB
11	ELECTRO-TECHNICAL-EMI/ EMC (Measure)	Antenna Symmetry	Using Vector Network Analyzer, Open Area Test Site (OATS) Reference Standard as per CISPR 16-1-6: 2014: Section 6.3.2 [CISPR 16-1-4:2019 Section 4.5.4.2] + Amd 1: 2017 + Amd 2: 2022	30 MHz to 300 MHz	2 dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11, SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2765 **Page No** 8 of 9

Validity 25/06/2022 to 24/06/2024 **Last Amended on** 30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
12	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Frequency / Generators	Using Rubidium Source & Frequency Counter by Direct Method	9 kHz to 14 GHz	1.3 Hz to 19.4 Hz
13	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Power / Generators (10 MHz to 18 GHz)	Using Power Sensors by Direct Method	10 dBm to (-) 40 dBm	0.41 dB to 0.64 dB
14	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Power / Generators (9 kHz to 10 MHz)	Using Power Sensors by Direct Method	13 dBm to (-) 20 dBm	0.36 dB to 0.41 dB
15	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Frequency / Receivers	Using Signal Generator, Reference Frequency Standard & Frequency Counter by Comparison / Reference Method	9 kHz to 18 GHz	1.3 Hz to 78 Hz



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

BNNSPEAG TEST & CALIBRATION LABORATORY INDIA PRIVATE LIMITED, 11/11,
SECTOR-3, RAJENDRA NAGAR, SAHIBABAD, GHAZIABAD, UTTAR PRADESH, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2765

Page No

9 of 9

Validity

25/06/2022 to 24/06/2024

Last Amended on

30/04/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
16	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Power / Receivers (9 kHz to 10 MHz)	Using Power Sensors with Signal Generators & Power Sensors by Comparison/ Substitution Method	10 dBm to (-) 20 dBm	0.43 dB to 0.47 dB
17	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Power/ Receivers (10 MHz to 18 GHz)	Using Power Sensors with Signal Generators & Power Sensors by Comparison/ Substitution Method	+10 dBm to (-) 40 dBm	0.47 dB to 0.70 dB

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.